

Journal homepage: http://www.journalijar.com

INTERNATIONAL JOURNAL OF ADVANCED RESEARCH

## **RESEARCH ARTICLE**

## Comparative Study of FIR Digital Filter for Noise Elimination in EMG Signal

Hemant kumar and Anjana Goen

Department of Electronics, RustamJI Institute of Technology BSF, Gwalior, India

#### Manuscript Info

#### Abstract

.....

Manuscript History:

Received: 14 October 2015 Final Accepted: 25 November 2015 Published Online: December 2015

Key words:

Electromyograme, Finite impulse response, Power spectral density & Mean squared error.

\*Corresponding Author

#### Hemant kumar

In Performance of EMG Signal, Signal acquisition must be noise free. This paper deals with the application of the digital FIR filter on the raw EMG signal. In this paper different window techniques used to design FIR filter and their Signal to noise ratio, Power spectral density, Mean & Standard deviation are compared.

.....

Copy Right, IJAR, 2015,. All rights reserved

# INTRODUCTION

In the example given below (Fig. 1), a raw surface EMG recording (EMG) was done for three static contractions of the biceps brachii muscle:

When the muscle is relaxed, a more or less noise-free EMG **Baseline** can be seen. The raw EMG baseline noise depends on many factors, especially the quality of the EMG amplifier, the environment noise and the quality of the given detection condition. Assuming a state-of-the-art amplifier performance and proper skin preparation, the averaged baseline noise should not be higher than 3-5 micro volts,

1 to 2 should be the target. The investigation of the EMG baseline quality is a very important checkpoint of every EMG measurement. Be careful not to interpret interfering noise or problems within the detection apparatus as "increased" base activity or muscle (hyper-) tonus.

The healthy relaxed muscle shows no significant EMG activity due to lack of depolarization and action potentials! By its nature, raw EMG spikes are of random shape, which means one raw recording burst cannot be precisely reproduced in exact shape. This is due to the fact that the actual set of recruited motor units constantly changes within the matrix/diameter of available motor units: If occasionally two or more motor units fire at the same time and they are located near the electrodes, they produce a strong superposition spike! By applying a smoothing algorithm (e.g. moving average) or selecting a proper amplitude parameter (e.g. area under the rectified curve), the non-reproducible contents of the signal is eliminated or at least minimized.

## **2. DIGITAL FIR FILTER**

The most essential tool of Digital signal processing is a digital filtering. A filter is a device when a signal is given; it changes to some desired form by changing its shape, amplitude, frequency or phase. They are usually employed to remove the noise, extract information signals and separate two or more combined signals. Digital filters are extremely used in noise cancellation, echo cancellation and also in the field of biomedical engineering to remove unwanted noise from EMG, ECG and EEG. The digital filters are divided into two basic types, Finite Impulse

Response (FIR) and Infinite Impulse Response (IIR) filters, which are known as non recursive and recursive filters. In the FIR system, the impulse response is of finite duration where as in the IIR system, the impulse response is of infinite duration. IIR filters are usually implemented using structures having feedback, that's why the present response of IIR filter is a function of present and past values of the excitation as well as the past value of the response. But the response of the FIR filter usually implemented using structures having no feedback so the response depends only on the present and past values of the input only.

#### A. DESIGN OF FIR LOW PASS FILTER

In this work different type of window techniques has been used for reduction of the noise and comparing SNR, PSD & standard deviation, these techniques are equi-ripple filter, Kaiser Window function, Hanning Window Function, Chebyshev Window Function, Hamming window function and Bartlett window function.

In this design the minimum order of the filter are selected and other parameters like Sampling frequency -4000 Hz, Pass band frequency 500Hz, Stop band frequency 1000 Hz,

Pass band attenuation 1dB, Stop band attenuation 80 dB,

Density factor 20. The response is shown below in figures.



Figure1-Normal EMG waveform.

## **3. METHODOLOGY**

In this paper noisy EMG signal are collected from three databases. The three examples of EMG data from: 1) a 44 year old man without history of neuromuscular disease; 2) a 62 year old man with chronic low back pain and neuropathy due to a right L5 radiculopathy; and 3) a 57 year old man with myopathy due to longstanding history of polymyosities, treated effectively with steroids and low-dose methotrexate. The data were recorded at 50 KHz and then down sampled to 4 KHz. During the recording process two analog filters were used: a 20 Hz high-pass filter and a 5K Hz low-pass filter. All signal are effected by noises and artifacts caused by some recourses. A low pass digital FIR filter designed for digitized signal at 4000 Hz per signal relative to real time [12].

The band pass-filtered signals were digitized at 4000 Hz per signal relative to real time using hardware constructed at the EMG. Three EMG data signals have a 40000 number of samples in 10 second. Sampling frequency of the data signal is 4000. Filter of noisy EMG data signal set up in two steps, in first step EMG input data signal extracted from the baseline drift and 10 db AWGN noise is introduced in input data signal as shown in figure 3. In the second step design a filter with the help of FDA tool in MAT LAB software and this noisy EMG data signal produce to as an input in designed filter with help of MATLAB coding. After the designing of filter, calculate the signal to noise ratio, power spectral density & standard deviation before and after filtering for window techniques, and compare these techniques on the basis of SNR, PSD & Standard deviation.

The test data results are compared in table -1 & table-2.



Figure 2-EMG signals with base line removal

## 4. RESULT

The results are calculated on EMG signals and accordingly, simulations are carried out on noisy EMG input and desired EMG signal. In this paper noise is reduced from different EMG data signals using FIR digital filter and SNR, PSD & standard deviation are calculated for various window techniques and compared these results before filtering and after filtering.

In this paper different EMG signals have been filtered by different FIR window techniques and their filtered values are used to calculate SNR, PSD & Standard Deviation. The aim of this paper is to find out the best window technique for different EMG signals and as we can see that in Table- 2, Bartlett window provides good results in terms of SNR(Signal to noise ratio), PSD(Power spectral density) & standard deviation for EMG signals.

Table 1:- SNRs, PSD & Standard deviation for EMG signal before filtering

EMG	SNR	PSD	Standard
Database			deviation
Healthy	8.2701	0.1318	10.8604
Myopathy	4.8785	0.1276	9.8134
Neuropathy	-9.1915	0.1135	13.2701

Table 2:-	SNRs.	PSD	& Standard	deviation f	or Different	EMG signals	s after filtering
I GOIC II	D1 1109		er Drumani a	actination	or Dimerent	DIVIO DIGINAL	, arear mooring

Window	EMG	SNR	PSD	Standard
Function	Database			deviation
Kaiser	Healthy	3.1943	0.0632	15.8604
window	Myopathy	3.5958	0.0558	14.9366
	Neuropaty	4.1155	0.0440	13.2701
Hanning	Healthy	2.7251	0.0704	16.7819
window	Myopathy	3.0993	0.0625	15.8141
	Neuropaty	3.5.68	0.0506	14.2330
Chebyshev	Healthy	2.8251	0.0688	16.5900
window	Myopathy	3.2250	0.0607	15.5870
	Neuropaty	3.6372	0.0491	14.0209
Bartlett	Healthy	2.5045	0.0741	17.2136
window	Myopathy	2.8324	0.0665	16.3075
	Neuropaty	3.2100	0.0542	14.7275
Hamming	Healthy	2.7300	0.0703	16.7725
window	Myopathy	3.1009	0.0665	15.8112
	Neuropaty	3.5075	0.0506	14.2318

	h total	A LINA	a hallala ha	nt land	in a state of the	probability,	Aparakal Manapal	uha /
THE PASS	Alle all and a state	In the second	Aller Malaland, In	Alburnate	de anne de la	An optimized in the	A. attended to	and a second
0	500	1000	1500	2000	2500	3000	3500	400
	heal	lthy signa	l after filte	ring from	Kaiser wi	ndow func	tion	
2	1				1		1	1
) •••	where the second	author has a series	and the second	and the second	Antopologica	free the state	name that have been	-
2	I	1000	1500	2000	7500	2000	2500	400
U	bealt	thy signal	after filter	ing from t	2000 Nanning W	indow fun	ction	400
	noun	anj orginar	unor meet	ing nom i	ranning a	indon idn	onon	-
) when	WHITH	where where where	-	initer and	-	Uninitally	-	-
2			1		1		1	and server
0	500	1000	1500	2000	2500	3000	3500	400
	health	y signal a	ifter filterir	ng from ch	iebyshev .	window fu	nction	
- Lu		an about					and and	
1 marine	All all and a feature of the	Mary an and the	HAR BALLE WAS	Andready	himmen	Provinsion of A	and a state of the	
0	500	1000	1500	2000	2500	3000	3600	400
0	heal	thy signal	after filte	ring from 1	Bartlett w	indow fun	ction	400
2								-
) when	-	with stranger		Serly sertent	No. of the second	-	-	and the state
$2^{\square}$	1		1	1	1		1	10/10/0000
0	500	1000	1500	2000	2500	3000	3500	400
	health	ny signal :	after filteri	ng from h	amming v	vindow fur	nction	
	and a k	an advers	and the second		-			and the second
	All and the second second	althout a set of the	HAR BALLE AND	Anthrough	hit to durate	Hardbard & F	Structure dist	No.

Figure 3- Noisy EMG healthy data signal and Filtered EMG healthy data signal using window Technique



Figure 4- Noisy EMG myopathy data signal and Filtered EMG myoopathy data signal using different windows technique



Figure 5- Noisy EMG neuropathy data signal and Filtered EMG neuropathy data signal using different windows technique

## 5. CONCLUSION & DISCUSSION

This paper is devoted to the problems and solutions on removal of Single Frequency Tones from Signals. The understanding of noise cancellation from EMG signal has been explained clearly using the techniques namely Kaiser, Hanning, Chebyshev, Hamming and Bartlett window techniques. It has been proposed a solution for high frequency noise interferences from original EMG signal, for Kaiser, Hanning, Chebyshev and Bartlett window function and results are compared on the basis of SNR, PSD & Standard deviation. So the Bartlett window has more noise reduction from noisy EMG data signal.

Table 1 and 2 has shown the SNR, PSD & Standard deviation before filtering and after filtering using FIR filter techniques. The results have been obtained which were required in purpose statement of the paper.

## ACKNOWLEDGMENT

I would like to thanks to all the experts who have contributed towards development of this paper

#### REFERENCES

- 1. Mohandas Choudhary, "ECG Signal Processing with FIR Digital Filter by Window Techniques", Vol. No. 02, 08, August 2012.
- Syed Zahurul Islam & Syed Zahidul Islam, "Performance Study of Adaptive Filtering Algorithms for Noise Cancellation of ECG Signal", *IEEE* 2009.
- 3. Wei Xing Zheng,"A Least- Squares based algorithm for FIR filtering with noisy data", IEEE 2003.

- 4. Van alste JA & Herrmann OE, "ECG baseline wander reduction using linear phase filters", compute .Biomed Res., 19(5);417-27; 1986Oct.
- 5. Lian.Y.Hoo, "Digital elliptical filter applications for noise reduction in ECG signal". *IEEE*, 2006.
- 6. Yong Lian and Poh Choo Ho,: "ECG noise reduction using multiplier-free FIR digital filters." IEEE, 2004.
- 7. Wei Xing Zheng, "A Least- Squares based algorithm for FIR filtering with noisy data", International Conference on Circuit and System, *IEEE*, Volume 4, May 2003.
- 8. Sara Moein, "FIR Cut-off Frequency Calculating for ECG Signal Noise Removing Using Artificial Neural Network", Verlag Berlin Heidelberg, pp. 124-131, 2010.
- 9. Syed Zahurul Islam & Syed Zahidul Islam, "Performance Study of Adaptive Filtering Algorithms for Noise Cancellation of ECG Signal", *IEEE* 2009.
- 10. Syed Zahurul Islam & Syed Zahidul Islam, "Performance Study of Adaptive Filtering Algorithms for Noise Cancellation of ECG Signal", *IEEE* 2009.
- 11. R. Acharya, & J.S. Suri, Advances in Cardiac Signal Processing, Springer, pp.1-50.
- 12. Sornmo, L. & Laguna, P.: Bioelectrical signal processing in cardiac neurological application. Elsevier, Amsterdam 2005.
- 13. "Comprarative study of chebyshev I and chebyshevII filter used for noise reduction in ECG signal".
- 14. http://en.wikipedia.org/wiki/Electrocardiography.
- 15. http://www.physionet.org/physiobank/database/PT.
- 16. G.M. Friesen & T.C. Jannett, "Acomparison of the noise sensitivity of nine QRS detection algorithm," *IEEETrans.Biomed.Eng*.Vol.37, pp.85–98,1990.
- 17. Mahesh S. & Agarwala S."Design and implementation of digital FIR filter equiripple notch filter on ECG signal for removal of power line interference". *Springer* 2008.
- 18. Youcef Ferdi,: "Improved Low pass Differentiator for Physiological Signal Processing", IEEE 2010.
- 19. Lian.Y.Hoo, "Digital elliptical filter application for noise reduction in ECG signal". IEEE 2006.
- 20. Yong Lian and Poh Choo Ho: "ECG noise reduction using multiplier-free FIR digital filters". IEEE 2004.
- 21. Moein, S.T., Advances in computation biology: "A MLP Neural Network for ECG Noise removed based on Kalman filter". Springer 2010.
- 22. In Haykin's Book, G.N. Swamy book on neural network.
- 23. Xin Liu\*& Yuan in Zheng, (*Member, IEEE*),: "Multiple Functional ECG Signal Processing for Wearable Applications of Long-Term Cardiac Monitoring". *IEEE* 2010.